

National Institute of Ayurveda, Jaipur

(Deemed University)

Ministry of AYUSH, Govt. of India

Jorawar Singh Gate, Old Amer Road, Jaipur (Rajasthan)-302002

QUOTATION INVITING NOTICE (QIN)

Sealed quotations are invited from the reputed and eligible interested Companies/Firms for Supply of **Swedan petika, Shirodhara Stand**, for **Panchakarma Department of National Institute of Ayurveda, Jaipur**, as per below specifications. Last date of receipt the quotation is **20 Aug 2026 (Date) Till 3.00 PM**.

S. No.	Description of Item	Detailed Specification	Quantity
1.	Swedan Peti with Digital Controller and Steam Generator	Primary Material: High-grade FRP (Fiber Reinforced Plastic) with food-grade ISO/NPG gel-coat finish for heat and moisture resistance. Dimensions & Design: Approx. 36" L × 30" W × 48" H head-out sitting cabinet with a non-porous exterior finish. Door & Latch System: Double front or single side-hinged door equipped with heavy-duty SS hinges and magnetic/rubber seals. Neck Opening: Anatomically contoured head cutout fitted with a medical-grade silicone gasket to prevent steam leakage. Seating System: 3-tier height-adjustable perforated FRP/SS stool to keep the patient's head safely outside the chamber. Base & Drainage: Sloped anti-skid moulded FRP floor with a 1/2" SS 304/brass discharge ball valve for sweat and condensate runoff. Internal Steam Piping: Perforated SS 304 or heavy-gauge copper distribution manifold targeting the lower limbs and torso uniformly. Mobility: 4 heavy-duty 360° swivel castor wheels with individual locking brakes. Digital Panel Display: Dual LED/LCD screen displaying real-time chamber temperature and setpoint temperature simultaneously. Temperature Control: Programmable 30°C to 60°C range with $\pm 1^\circ\text{C}$ precision using a waterproof PT100 thermistor probe. Digital Timer: 1 to 60 minutes programmable countdown timer with automatic steam cut-off relay and alarm chime. Electrical Safety: Built-in ELCB (Earth Leakage Circuit Breaker), MCB overload switch, and high-temperature thermal safety cut-off at 65°C. Low-Water Protection: Automatic water-level sensor relay that cuts power to boiler heating elements if water falls below the minimum limit. Steam Generator Boiler: Heavy-gauge SS 304 / Aluminium boiler vessel with 10 to 20 Liters total water holding capacity. Heating Element: 2.0 kW to 3.0 kW ISI-marked Stainless Steel immersion heater operating on 220V–240V AC single-phase power. Herbal Infusion Basket: Removable SS 304 perforated herbal basket inside the boiler line for direct Kwatha/Dravyainfusion. Pressure Safety: Mechanical spring-loaded brass pressure relief valve and an adjustable brass steam flow gate valve. Delivery Line & Gauge: Braid-reinforced insulated flexible silicone/Teflon steam hose with an external visual water-level sight tube.	5
2.	Shirodhara Stand	Sturdy steel shirodhara stand. Made in heavy duty stainless steel pipes (SS 304). Ballasted for perfect balance & stability. The shirodhara vessel can be positioned anywhere in both horizontal & vertical directions. Oil Vessel & holder provided. Head rest Model is preferred.	4

General Terms & Conditions :

1. The Supplier will have to complete the supply at **Panchakarma Department of National Institute of Ayurveda, Jaipur within 7/15/30 Days from supply order date.**
2. The Quotation Price shall be in Rupees including all Taxes and no extra prices such as transportation expenses, Installation etc. are payable by the Institute.
3. After due evaluation of the quotation(s), Institute will award the Supply Order to the lowest evaluated responsive firm.
4. Payment shall be made only after completion of the Supply, Installation and submission of bill which will be subject to satisfactory report (certified by the Head of Department or Committee constituted for the purpose).
5. **The Vice Chancellor, NIA Jaipur** shall be the final authority to reject full or any part of the order which is not confirming to the requirements and other terms and conditions.
6. No payment shall be made for the unsatisfactory work/Item. Item/Machine will be under comprehensive **warranty for One Year (s) from the date of installation.**
7. GST certificate must be enclosed by the firm. Quotation will not be entertained without GST certificate.
8. The Firm must enclose **a Certificate on Firm's Letter Head duly Sealed and Signed** along with the Quotation stating that "It is certified that our firm M/s.....(complete address of firm) is competent to supply (Name of Machine/Item) and not debarred by Department of Commerce or Ministry/Department concerned". **In absence of this certificate, Quotation will not be entertained.**
9. The quotations in sealed envelope, super scribing clearly on the envelope "**Quotations for Supply of Swedan petika, Shirodhara Stand for Panchakarma Department of National Institute of Ayurveda, Jaipur** should reach **NIA, Jaipur** on or before **20 Aug 2026 Date)** at the below mentioned address :-

The Vice Chancellor, National Institute of Ayurveda (Deemed University), Madhav Vilas Palace, Jorawar Singh Gate, Amer Road, Jaipur - 302002

Sd/-

VICE CHANCELLOR